






Fusion QbD

***Advanced QbD Software for
Analytical Method Validation
and Transfer***

A Complete Solution for APLM Stages 1 and 2

Analytical Procedure Lifecycle Management Workflow





Printed on: Thu Mar 07 2024, 04:42:23 PM(EST) Status: Currently Official on 07-Mar-2024 DocId: GUID-35D7E47E-65E5-49B7-B4CC-4D96FA230821_2_en-US
 Printed by: George Cooney Official Date: Official as of 01-May-2022 Document Type: GENERAL CHAPTER @2024 USPC
 Do Not Distribute DOI Ref: 468ba DOI: https://doi.org/10.31003/USPNF_M10975_02_01 1

Add the following:
 ▲(1220) ANALYTICAL PROCEDURE LIFE CYCLE

INTRODUCTION
 This general chapter holistically considers the validation activities that take place across the entire life cycle of an analytical procedure and provides a framework for the implementation of the life cycle approach. The analytical procedure life cycle approach described here is consistent with the quality by design concepts described in International Council for Harmonisation (ICH) guidelines. The procedure life cycle approach emphasizes the importance of sound scientific approaches and quality risk management for the development, control, establishment, and use of analytical procedures. Total error is used in this chapter; however, measurement uncertainty can also be used. The procedure life cycle approach is applicable to all types of analytical procedures, and the extent of effort should be consistent with the complexity of the procedure and the criticality of the quality attribute to be measured. The life cycle approach can be considered optional, and any of the elements can be applied on the basis of how the procedure is used. Elements of the life cycle approach can also be applied retrospectively if deemed useful or in early stages of development with the appropriate modifications. Elements of life cycle management of analytical procedures are also discussed in *Analytical Procedures and Methods Validation for Drugs and Biologics* (Guidance for Industry, FDA 2015). Validation of an analytical procedure is the process by which it is established, through laboratory studies, that the performance of the procedure meets the requirements for the intended analytical applications. Validation, or demonstration that a procedure is suitable for the intended purpose, takes place during the entire procedure life cycle, beginning during the initial procedure design activities and extending through routine use. These activities include the formal procedure validation, verification, and transfer of procedures, as well as establishing and assuring adherence to an appropriate set of procedure controls and system suitability requirements. The procedure life cycle is comprised of the analytical target profile (ATP) and three stages, which are introduced below and shown in Figure 1. The ATP defines the criteria for the procedure performance characteristics that are linked to the intended analytical application and the quality attribute to be measured. It applies to all stages of the procedure life cycle. For quantitative procedures, the ATP describes the required quality of the reportable value since the reportable value generated using a qualified analytical procedure provides the basis for key decisions regarding compliance of a test article with regulatory, compendial, and manufacturing limits. The acceptable level of risk of making an incorrect decision can also be considered when establishing ATP criteria. **Stage 1:** Procedure design encompasses procedure development, which consists of the analytical technology and sample preparation. It includes understanding gained through knowledge gathering, systematic procedure development experiments, and risk assessments and associated lab experiments. The output of Stage 1 includes:
 1. A set of procedure conditions that minimizes procedure bias to a suitable level, can provide acceptable precision, and can meet the ATP criteria
 2. An understanding of the effect of procedure parameters (e.g., temperature, wavelength, flow rate, etc.) on procedure performance
 3. Optimization of performance characteristics of the analytical procedure such as accuracy, precision, the appropriateness of any calibration model, specificity and limit of quantitation (as far as applicable); this includes a preliminary replication strategy for samples and standards
 4. An initial analytical control strategy (ACS), which is a set of controls (system suitability tests [SSTs] and other procedure-specific controls) needed to ensure proper performance
Stage 2: Procedure performance qualification consists of studies designed to demonstrate that the procedure is suitable for its intended purpose. This involves confirmation that the reportable values generated by application of the analytical procedure meet the ATP criteria as well as confirmation of procedure performance characteristics through the traditional validation, verification, or transfer studies. Data generated during Stage 1 can be used if available and suitable. At the end of Stage 2, the replication strategy and the performance of the procedure is confirmed to meet the ATP and other criteria.
Stage 3: Ongoing procedure performance verification involves monitoring the analytical procedure during routine use and confirming that the performance continues to meet ATP criteria. Monitoring ensures that the performance of the procedure is maintained at an acceptable level over the procedure lifetime. It can also provide an early indication of potential performance issues or adverse trends and aid in identifying required changes for the analytical procedure. Confirming procedure performance after changes ensures that the modified procedure will produce reportable values that meet the criteria defined in the ATP. More details about the procedure life cycle are described in the subsequent sections.

https://online.uspnc.com/uspnfdocument/1_GUID-35D7E47E-65E5-49B7-B4CC-4D96FA230821_2_en-US 1/12

Printed on: Sat Mar 16 2024, 07:59:46 PM(EST) Status: Currently Official on 17-Mar-2024 DocId: GUID-13ED4BEB-4086-43B5-A7D7-994A02AF25C8_7_en-US
 Printed by: George Cooney Official Date: Official as of 01-May-2018 Document Type: GENERAL CHAPTER @2024 USPC
 Do Not Distribute DOI Ref: safim DOI: https://doi.org/10.31003/USPNF_M88646_07_01 1

Add the following:
 ▲(1210) STATISTICAL TOOLS FOR PROCEDURE VALIDATION

1. INTRODUCTION
2. CONSIDERATIONS PRIOR TO VALIDATION
3. ACCURACY AND PRECISION
 - 3.1 Methods for Estimating Accuracy and Precision
 - 3.2 Combined Validation of Accuracy and Precision
4. LIMITS OF DETECTION AND QUANTITATION
 - 4.1 Estimation of LOD
 - 4.2 Estimation of LOQ
5. CONCLUDING REMARKS

REFERENCES

1. INTRODUCTION

This chapter describes utilization of statistical approaches in procedure validation as described in *Validation of Compendial Procedures* (1225). For the purposes of this chapter, "procedure validation" refers to the analytical procedure qualification stage of the method life cycle, following design and development and prior to testing. Chapter (1225) explains that capabilities of an analytical procedure must be validated based on the intended use of the analytical procedure. Chapter (1225) also describes common types of uses and suggests procedure categories (I, II, III, or IV) based on the collection of performance parameters appropriate for these uses. Performance parameters that may need to be established during validation include accuracy, precision, specificity, detection limit [limit of detection, (LOD)], quantitation limit, linearity, and range. In some situations (e.g., biological assay), relative accuracy takes the place of accuracy. This chapter focuses on how to establish analytical performance characteristics of accuracy, precision, and LOD. For quantitative analytical procedures, accuracy can only be assessed if a true or accepted reference value is available. In some cases, it will be necessary to assess relative accuracy. In many analytical procedures, precision can be assessed even if accuracy cannot be assessed. The section addressing LOD can be applied to limit tests in Category II. The other analytical performance characteristics noted in (1225), which include specificity, robustness, and linearity, are out of scope for this chapter. Because validation must provide evidence of a procedure's fitness for use, the statistical hypothesis testing paradigm is commonly used to conduct validation consistent with (1225). Although some statistical interval examples are provided in 3. *Accuracy and Precision*, these methods are not intended to represent the only approach for data analysis, nor to imply that alternative methods are inadequate. Table 1 provides terminology used to describe an analytical procedure in this chapter. The definitions for individual determination and reportable value are in alignment with *General Notices, 7.10 Interpretation of Requirements*.

Table 1. Analytical Procedure Validation Terminology

Terminology	Description
Laboratory sample	The material received by the laboratory
Analytical sample	Material created by any physical manipulation of the laboratory sample, such as crushing or grinding
Test portion	The quantity (aliquot) of material taken from the analytical sample for testing
Test solution	The solution resulting from chemical manipulation of the test portion such as chemical derivatization of the analyte in the test portion or dissolution of the test portion
Individual determination (ID)	The measured numerical value from a single unit of test solution
Reportable value	Average value of readings from one or more units of a test solution

Not all analytical procedures have all stages shown in Table 1. For example, liquid laboratory samples that require no further manipulations immediately progress to the test solution stage. Demonstration that a reportable value is fit for a particular use is the focus of analytical validation. Table 2 provides an example of the Table 1 terminology for a solid oral dosage form.

Table 2. Example for Coated Tablets

Terminology	Description	
Laboratory sample	100 coated tablets	
Analytical sample	20 tablets are removed from the laboratory sample and are crushed in a mortar and pestle	
Test portion	Replicate 1: 1 g of crushed powder aliquot from the analytical sample	Replicate 2: 1 g of crushed powder aliquot from the analytical sample

https://online.uspnc.com/uspnfdocument/1_GUID-13ED4BEB-4086-43B5-A7D7-994A02AF25C8_7_en-US 1/12



METHOD VALIDATION MODULE

- Full Validation Experiment Suite
- Instant Analysis and Reporting
- Advanced Method Transfer Support
- Meets all Regulatory Requirements

Critical QbD Capability

FMV

Supports All Install Environments (Citrix Ready Certified)



Full 21 CFR Part 11 Compliance Support



Complete Method Validation Experiment Suite



Simple Experiment Workflows



Full LC Experiment Automation



USP 1210> Tolerance and Prediction Interval Metrics



- Replication Strategy and Total Analytical Error
- Accuracy and Repeatability
- Analytical Method Transfer

Critical QbD Capability

FMV

Supports All Install Environments (Citrix Ready Certified)



Full 21 CFR Part 11 Compliance Support



Complete Method Validation Experiment Suite



Simple Experiment Workflows



Full LC Experiment Automation



USP 1210> Tolerance and Prediction Interval Metrics



- Replication Strategy and Total Analytical Error
- Accuracy and Repeatability
- Analytical Method Transfer

Supports All Install Environments

Install Environment

FMV

Standalone (Workstation)



Network (Enterprise)



Citrix Ready Certified




Fully Qualifiable for GXP Environments*



* – Fusion QbD is operating in the GxP environments of international pharmaceutical companies worldwide.

Example Network Deployment



Critical QbD Capability

FMV

Supports All Install Environments (Citrix Ready Certified)



Full 21 CFR Part 11 Compliance Support



Complete Method Validation Experiment Suite



Simple Experiment Workflows



Full LC Experiment Automation



USP 1210> Tolerance and Prediction Interval Metrics



- Replication Strategy and Total Analytical Error
- Accuracy and Repeatability
- Analytical Method Transfer

Full Support for 21 CFR 11 Compliance

FMV

Full integration of all e-record and all e-signature features and functions required to support full 21 CFR 11 compliance.



Integrated Workflow Management and Secure Project Management Systems.



Full audit trail, including bi-directional auditing of all data exchanges with the CDS.



Why Audit Trail is Important!

Where did this data come from?
Empower Project?
Results Set?
Chromatograms?



Who imported this data – was the data modified?

Audit Log Filter Options

Date

Enable

Starting Date:

March 2020						
Sun	Mon	Tue	Wed	Thu	Fri	Sat
23	24	25	26	27	28	29
1	2	3	4	5	6	7
8	9	10	11	12	13	14
15	16	17	18	19	20	21
22	23	24	25	26	27	28
29	30	31	1	2	3	4

Ending Date:

March 2020						
Sun	Mon	Tue	Wed	Thu	Fri	Sat
23	24	25	26	27	28	29
1	2	3	4	5	6	7
8	9	10	11	12	13	14
15	16	17	18	19	20	21
22	23	24	25	26	27	28
29	30	31	1	2	3	4

Users

Enable

Available:

Administrator

Included:

Events

Enable

Available:

- Print Reports
- Experiment Setup
- Enable User Defined Option
- Generate Design
- Export Experiment Design
- Export Testing Design
- Matrix Master Wizard
- Edit Run No. Labels
- Robustness Simulator
- Create Testing Design
- Delete Testing Design
- Response Reductions

Included:

- Import Responses
- Create/Edit Response Data

OK Cancel ?

Critical QbD Capability

FMV

Supports All Install Environments (Citrix Ready Certified)



Full 21 CFR Part 11 Compliance Support



Complete Method Validation Experiment Suite



Simple Experiment Workflows




Full LC Experiment Automation



USP 1210> Tolerance and Prediction Interval Metrics



- Replication Strategy and Total Analytical Error
- Accuracy and Repeatability
- Analytical Method Transfer



- Replication Strategy*
 - Specificity
 - Filter Validation
 - Sample Solution Stability
 - Accuracy*
 - Linearity & Range
 - Repeatability*
 - Accuracy / Linearity / Repeatability*
[Combined as per ICH Q2(R1)]
 - LOQ*, LOD*
 - Intermediate Precision and Reproducibility
 - Validation Robustness – LC
 - Validation Robustness – Non-LC
[e.g. Sample Preparation, Dissolution]
 - Method Transfer Study Support*
- * – integration of USP <1210> Tolerance & Prediction Intervals]

Critical QbD Capability

FMV

Supports All Install Environments (Citrix Ready Certified)



Full 21 CFR Part 11 Compliance Support



Complete Method Validation Experiment Suite



Simple Experiment Workflows



Full LC Experiment Automation



USP 1210> Tolerance and Prediction Interval Metrics



- Replication Strategy and Total Analytical Error
- Accuracy and Repeatability
- Analytical Method Transfer

Example: Accuracy / Linearity / Repeatability – Combined Experiment

Create New Work File

Project

Select Project: Audit Logging Enabled

Project Name:

Instrument

Instrument Type: LC
Data System: Waters Empower
Pump Module: Quaternary


Sample Compound Type

Small Molecule Large Molecule

Experiment Phase

Experiment Type

- Analytical Capability
- Specificity
- Accuracy
- Linearity and Range
- Repeatability
- Accuracy, Linearity, Repeatability**
- Robustness
- Robustness Non-LC



1. Simple Experiment Setup Template

Experiment Setup

Replication Settings

Global Compound Settings

Assay Type

No. of Compounds

No. of Levels per Compound

100% Std. Level

Compound Name	Units		Level Settings
Compound 1	%	<input type="button" value="-0.00"/> <input type="button" value="+0.00"/>	Level 1 <input style="width: 50px;" type="text" value="80"/> Level 2 <input style="width: 50px;" type="text" value="90"/> Level 3 <input style="width: 50px;" type="text" value="100"/> Level 4 <input style="width: 50px;" type="text" value="110"/> Level 5 <input style="width: 50px;" type="text" value="120"/>

Compound Name	Units		Level Settings
Compound 2	%	<input type="button" value="-0.00"/> <input type="button" value="+0.00"/>	Level 1 <input style="width: 50px;" type="text" value="80"/> Level 2 <input style="width: 50px;" type="text" value="90"/> Level 3 <input style="width: 50px;" type="text" value="100"/> Level 4 <input style="width: 50px;" type="text" value="110"/> Level 5 <input style="width: 50px;" type="text" value="120"/>

Create and Maintain Templates.

Set Automatic E-Review and E-Approve Loops.

2. Standards Setup Options

Standards Setup

Standards Strategy
 Calibration and Check Standards
 <None Selected>
 Bracketing - NonOverlap
 Grand Average
 Calibration and Check Standards
 Multi-level Bracketing - Overlap

No. of Repeat Injections per Level 1

Check Standards Scheme
 No. of Standards per Group 1
 No. of Injections Between Groups 5

Experiment Design


	Run No.	API	Impurity 1	Impurity 2	Degradan t A	Degradan t B
1	CAL - L1.1	---	---	---	---	---
2	CAL - L2.1	---	---	---	---	---
3	CAL - L3.1	---	---	---	---	---
4	CAL - L4.1	---	---	---	---	---
5	CAL - L5.1	---	---	---	---	---
6	Chk - 1.a	100	100	100	100	100
7	1.a	80	80	80	80	80
8	1.b	80	80	80	80	80
9	1.c	80	80	80	80	80
10	1.d	80	80	80	80	80
11	1.e	80	80	80	80	80
12	Chk - 1.b	100	100	100	100	100
13	2.a	90	90	90	90	90

Validation Status: Your settings are valid.

Clear Reset Next >> Cancel ?

Flexible setup of the required Standards Strategies.

3. Auto-generated Experiment Design



Analytical Method Validation - Untitled1.smae

File Edit Activity Tools Window Help

Design of Experiments

- Create a Design
- **Design Reports**

Data Entry / Analysis

- Data Entry
- Data Analysis

Reporting Toolkit

- Fusion Reporter
- Audit Log Reporter

Design Reports

- Experiment Design
- Instrument Report
- Experiment Design**
- Experiment Setup

Name: Administrator
Company: S-Matrix
Project: User Defined
Date: August 19, 2016 1:09:19 PM PDT [GMT-07:00]

Experiment Design - Accuracy/Linearity/Repeatability

Experiment Design Matrix

Run No.	API (%)	Impurity 1 (%)	Impurity 2 (%)	Degradant A (%)	Degradant B (%)
CAL - L1.1	100	100	100	100	100
CAL - L2.1	100	100	100	100	100
CAL - L3.1	100	100	100	100	100
CAL - L4.1	100	100	100	100	100
CAL - L5.1	100	100	100	100	100
Chk - 1.a	100	100	100	100	100
1.a	80	80	80	80	80
1.b	80	80	80	80	80
1.c	80	80	80	80	80
2.a	90	90	90	90	90

Ready | modified

Comprehensive, automated reporting.

4. Analysis Wizard for CDS Imported Results

Method Validation - Small Molecule Data Analysis

Accuracy | **Linearity** | Repeatability

Select Response for Analysis
Peak Area

API

Perform Data Analysis

Compound-based Acceptance Criteria

Linearity Regression ($r \geq$)

Intercept % Bias Calculation Options

Data Based Model Based

Intercept |% Bias| \leq

LOQ / LOD

Calculation Method(s)

ICH-Q2B USP <1210>

Significance Level %

LOQ Calculation(s)

Residual Std. Dev. Intercept Std. Dev.

LOD Calculation(s)

Residual Std. Dev. Intercept Std. Dev.

Level-based Acceptance Criteria

Computed Results

Include Response Factor

Level	Linearity % Bias of Residuals \leq	Response Factor % Bias \leq
1.000	5.00	
2.000	5.00	
4.000	5.00	
5.000	5.00	

Source Data

Level	Individual Results Spec. Lower Limit	Individual Results Spec. Upper Limit
1.000	1627663	1798996
2.000	3340993	3512326
4.000	6767652	6938985
5.000	8480023	8652315

The settings are valid.

Back Finish Cancel

Associate responses with Analyses:
For Example – Amount for Accuracy and Area for Linearity.



Set Global and Level-specific Acceptance Criteria, including Spec Limits for Data

Include LOQ and LOD. Select Calculation Method Options.

5. Instant Analysis, Graphing, and Reporting

ICH Q2(R2):

Data derived from the regression line may help to provide mathematical estimates of the linearity. A plot of the data, the correlation coefficient or coefficient of determination, y-intercept and slope of the regression line should be provided. An analysis of the deviation of the actual data points from the regression line is helpful for evaluating linearity (e.g., for a linear response, the impact of any non-random pattern in the residuals plot from the regression analysis should be assessed).




Fusion QbD instantly creates formal reports with all required tables and graphs.

5. Instant Analysis, Graphing, and Reporting

ICH Q2(R2):



Representative data (e.g., chromatograms, electropherograms, spectra, biological response) should be used to demonstrate specificity and relevant components should be labelled, if appropriate.

For a purity or impurity test, discrimination can be established by stressing or spiking product to achieve appropriate levels of impurities or related substances and demonstrating the absence of interference.



Reports can be augmented with images of relevant chromatograms.

5. Instant Analysis, Graphing, and Reporting



Reports meet all output format requirements:

.TXT / .RTF / .DOC / .PDF / .HTML / XLSX

Critical QbD Capability

FMV

Supports All Install Environments (Citrix Ready Certified)



Full 21 CFR Part 11 Compliance Support



Complete Method Validation Experiment Suite



Simple Experiment Workflows



Full LC Experiment Automation




USP 1210> Tolerance and Prediction Interval Metrics



- Replication Strategy and Total Analytical Error
- Accuracy and Repeatability
- Analytical Method Transfer

Automated Experiment Workflow

Steps 1 and 2



Generates Selected
Validation Experiment

Automatically Builds
Sequence with
Standards Protocol and
Assigns Method


Step 3

Chromatography Data Software (CDS)


Vial	Inj Vol (uL)	# of Injs	Label	SampleName	Level	Function	Method Set / Report Method	Label Reference	Processing	Run Time (Minutes)	Data Start (Minutes)	Next Inj. Delay (Minutes)	Column Position	Auto Additions	SampleWeight	Dilution
1						Condition Column				6.70	0.00	0.00	No Change			
2						Condition Column	Text Mix pH 001_017			0.10	0.00	0.00	No Change			
3						Equilibrate	Text Mix pH 001_017			3.00	0.00	7.95	No Change			
4	1	2.0	1	Unk-000-000	Blank - 1	Inject Samples	Text Mix pH 001_017		Normal	10.50	0.00	1.50			1.00000	1.00000
5						Condition Column	Text Mix pH 001_001			0.10	0.00	0.00	No Change			
6						Equilibrate	Text Mix pH 001_001			3.00	0.00	0.00	No Change			
7	2	2.0	1	Unk-001-001	1.a.1.a	Inject Samples	Text Mix pH 001_001		Normal	10.50	0.00	1.50			1.00000	1.00000
8						Condition Column	Text Mix pH 001_002			0.10	0.00	0.00	No Change			
9						Equilibrate	Text Mix pH 001_002			3.00	0.00	0.00	No Change			
10	2	2.0	1	Unk-001-002	2.a.1.a	Inject Samples	Text Mix pH 001_002		Normal	10.50	0.00	1.50			1.00000	1.00000
11						Condition Column				6.70	0.00	0.00	No Change			
12						Condition Column	Text Mix pH 001_003			0.10	0.00	0.00	No Change			
13						Equilibrate	Text Mix pH 001_003			3.00	0.00	0.00	No Change			
14	2	2.0	1	Unk-001-003	3.a.1.a	Inject Samples	Text Mix pH 001_003		Normal	10.50	0.00	1.50			1.00000	1.00000
15						Condition Column				6.70	0.00	0.00	No Change			
16						Condition Column	Text Mix pH 001_004			0.10	0.00	0.00	No Change			
17						Equilibrate	Text Mix pH 001_004			3.00	0.00	0.00	No Change			
18	2	2.0	1	Unk-001-004	4.a.1.a	Inject Samples	Text Mix pH 001_004		Normal	10.50	0.00	1.50			1.00000	1.00000
19						Condition Column	Text Mix pH 001_005			0.10	0.00	0.00	No Change			
20						Equilibrate	Text Mix pH 001_005			3.00	0.00	0.00	No Change			
21	2	2.0	1	Unk-001-005	5.a.1.a	Inject Samples	Text Mix pH 001_005		Normal	10.50	0.00	1.50			1.00000	1.00000
22						Condition Column				6.70	0.00	0.00	No Change			
23						Condition Column	Text Mix pH 001_006			0.10	0.00	0.00	No Change			
24						Equilibrate	Text Mix pH 001_006			3.00	0.00	0.00	No Change			

Automated, Audited Data Exchange
Preserves Data Integrity


Automated Experiment Workflow



Chromatography Data Software (CDS)



Run No	Vial No	# of Pkts	Label	Sample Name	Level	Function	Method Set / Report Defect	Label Reference	Processing	Run Time (Minutes)	Data Start (Minutes)	Inject MS Delay (Minutes)	Column Position	Auto Adjusts	Sample Weight	Date
1						Condition Column				8.70	0.00	0.00	No Change			
2						Condition Column	Test file pH 01_017			8.18	0.00	0.00	No Change			
3						Equilibrate	Test file pH 01_017			3.00	0.00	7.90	No Change			
4	1	2.0	1	UW-005-000	Blank - 1	Inject Samples	Test file pH 01_017		Normal	10.00	0.00	1.00			1.00000	1.00000
5						Condition Column	Test file pH 01_021			8.10	0.00	0.00	No Change			
6						Equilibrate	Test file pH 01_021			3.00	0.00	0.00	No Change			
7	2	2.0	1	UW-001-001	1 x 1 x	Inject Samples	Test file pH 01_021		Normal	10.00	0.00	1.00			1.00000	1.00000
8						Condition Column	Test file pH 01_022			8.18	0.00	0.00	No Change			
9						Equilibrate	Test file pH 01_022			3.00	0.00	0.00	No Change			
10	2	2.0	1	UW-001-002	2 x 1 x	Inject Samples	Test file pH 01_022		Normal	10.00	0.00	1.00			1.00000	1.00000
11						Condition Column	Test file pH 01_023			8.70	0.00	0.00	No Change			
12						Equilibrate	Test file pH 01_023			8.18	0.00	0.00	No Change			
13						Equilibrate	Test file pH 01_023			3.00	0.00	0.00	No Change			
14	2	2.0	1	UW-001-003	2 x 1 x	Inject Samples	Test file pH 01_023		Normal	10.00	0.00	1.00			1.00000	1.00000
15						Condition Column	Test file pH 01_024			8.70	0.00	0.00	No Change			
16						Equilibrate	Test file pH 01_024			8.18	0.00	0.00	No Change			
17	2	2.0	1	UW-001-004	1 x 1 x	Inject Samples	Test file pH 01_024		Normal	10.00	0.00	1.00			1.00000	1.00000
18						Condition Column	Test file pH 01_025			8.18	0.00	0.00	No Change			
19						Equilibrate	Test file pH 01_025			3.00	0.00	0.00	No Change			
20	2	2.0	1	UW-001-005	1 x 1 x	Inject Samples	Test file pH 01_025		Normal	10.00	0.00	1.00			1.00000	1.00000
21						Condition Column	Test file pH 01_026			8.70	0.00	0.00	No Change			
22						Condition Column	Test file pH 01_026			8.18	0.00	0.00	No Change			
23						Equilibrate	Test file pH 01_026			3.00	0.00	0.00	No Change			



Automated, Audited Data Exchange Preserves Data Integrity

Run on Your LC System





- ✓ Solvent Selection Valves
- ✓ Column Switching Valves

Alliance HPLC



Alliance iS HPLC



Acquity Binary



Acquity H-Class



Acquity Arc



Acquity UPC²





OpenLab –
ChemStation
Edition



Solvent Selection Valves



Column Switching Valves

Agilent 1100s
And 1200s



Agilent 1260
Infinity Series



Agilent 1260
Infinity II Series



Agilent 1290
Infinity Series



Agilent 1290
Infinity II Series





- ✓ Solvent Selection Valves
- ✓ Column Switching Valves

UltiMate LCs



Vanquish Horizon And Flex LCs



Critical QbD Capability

FMV

Supports All Install Environments (Citrix Ready Certified)



Full 21 CFR Part 11 Compliance Support



Complete Method Validation Experiment Suite



Simple Experiment Workflows




Full LC Experiment Automation



USP 1210> Tolerance and Prediction Interval Metrics



- **Replication Strategy and Total Analytical Error**
- **Accuracy and Repeatability**
- **Analytical Method Transfer**



2. CONSIDERATIONS PRIOR TO VALIDATION

How many individual determinations will compose the reportable value, and how will they be aggregated?

- To answer this question, it is necessary to understand the contributors to the procedure variance and the ultimate purpose of the procedure.

Estimation of variance components during pre-validation provides useful information for making this decision.

Critical QbD Capability

FMV

Supports All Install Environments (Citrix Ready Certified)



Full 21 CFR Part 11 Compliance Support



Complete Method Validation Experiment Suite



Simple Experiment Workflows



Full LC Experiment Automation



USP 1210> Tolerance and Prediction Interval Metrics



- **Replication Strategy and Total Analytical Error**
- Accuracy and Repeatability
- Analytical Method Transfer

Replication Strategy Optimization



ICH Q14

Reportable Result: the result as generated by the analytical procedure after calculation or processing and applying the described sample replication. (ICH Q2)

ICH Q2(R2)

The experimental design of the validation study should reflect the number of replicates used in routine analysis to generate a reportable result.

USP <1220>

Stage 1:

Optimization of performance characteristics of the analytical procedure such as accuracy, precision, ...; this includes a preliminary replication strategy for samples and standards.

Replication Strategy Experiment

Define your Proposed Replication Strategy, Target Result Value, Acceptance Limits, Desired TAE Limits, and your Desired Probability and Tolerance (Confidence Interval).

Replication Strategy Analysis Setup

REFERENCE TABLE

Sigma	FPT	CP
1.00	322.17	0.33
2.00	44.43	0.67
3.00	2.7	1.00
4.00	0.07	1.33
5.00	0.00054	1.67
6.00	0.000002	2.00


Replication Strategy
Number of preparations x Number of injections per preparation. Together these settings define the data which are averaged into the reportable value for the documented method.

Total Analytical Error (TAE) Limits
The \pm value = the minimum allowable \pm distance of a given reportable value from the corresponding acceptance limit.

Select a Replication Strategy
No. of preparation replicates per sample: No. of injections per preparation replicate:

Enabled	Responses	Target Value	\pm Acceptance Limits	\pm Total Analytical Error (TAE) Limits	TAE # σ Width	Interval Type	Desired Probability (%)	Tolerance Alpha (%)
<input checked="" type="checkbox"/>	API - Amount	100.000	2.000	0.300	2 σ	Tolerance	95.00	5.00

The settings are valid.



The diagram illustrates the Total Analytical Error (TAE) and Guard Bands. It shows a central 'Reportable Value' with a 'Safe and Efficacious Range' indicated by a blue double-headed arrow. On either side of this range are 'Guard Bands', followed by 'Acceptance Zones (Meets Specification)' and 'Rejection Zones'. Two normal distribution curves are shown, one for the 'Lower Acceptance Limit' and one for the 'Upper Acceptance Limit', with their respective -2σ and $+2\sigma$ points marked. A smaller 'Total Analytical Error Distribution' curve is also shown in the top right corner.

Replication Strategy for the Reportable Value

Between Variables Components of Variation

Variable Name	Variance	Standard Deviation	Degrees of Freedom	t-table Value	(+/-) 95% Confidence Limits	Error Contribution (%)
Sample Preparation	0.008	0.092	4	2.7764	0.25	94.17
Injection	0.001	0.023	20	2.0860	0.04	5.83

TOST Analysis Results Summary

Statistic	Value	Pass/Fail
TAE Width (2σ) - Target	± 0.300	
Computed TAE Width (2σ)	± 0.156	Pass
FPT	<0.0001	
Cp	12.2271	
Variance	0.003	
Standard Deviation	0.055	
% RSD	0.05	
% CV	0.05	

Tolerance Interval Analysis Results


Interval Setting	Value	Number of Preparations	Number of Injections per Preparation
Target	100.000	3	1
Acceptance Limits	± 2.000		
Desired Probability %	95.00		
Tolerance Alpha %	5.00		
Grand Mean	100.051		
Computed Tolerance Interval	± 0.293	Pass	
Required Guard Band Width	± 0.300		

The computed Tolerance Interval falls within the defined Total Analytical Error Limits.

Overall Error in a Single Determination

Statistic	Value
Mean	100.051
Variance	0.009
Standard Deviation	0.094
% RSD	0.094

No. of Injections	No. of Preparations										
	1	2	3	4	5	6	7	8	9	10	
1	$\pm 2\sigma$	0.2710	0.191	0.1564	0.1355	0.1212	0.1106	0.1024	0.0958	0.0903	0.0857
	T.I.	0.6210	0.380	0.2927	0.2455	0.2151	0.1936	0.1774	0.1647	0.1543	0.1456
2	$\pm 2\sigma$	0.2670	0.1888	0.1541	0.1335	0.1194	0.1090	0.1009	0.0944	0.0890	0.0844
	T.I.	0.5299	0.3421	0.2698	0.2295	0.2029	0.1838	0.1693	0.1577	0.1482	0.1402
3	$\pm 2\sigma$	0.2657	0.1878	0.1534	0.1328	0.1188	0.1085	0.1004	0.0939	0.0886	0.0840
	T.I.	0.4971	0.3288	0.2620	0.2240	0.1988	0.1806	0.1665	0.1553	0.1461	0.1384
4	$\pm 2\sigma$	0.2650	0.1874	0.1530	0.1325	0.1185	0.1082	0.1002	0.0937	0.0883	0.0838
	T.I.	0.4801	0.3221	0.2580	0.2213	0.1968	0.1789	0.1652	0.1542	0.1451	0.1375
5	$\pm 2\sigma$	0.2646	0.1871	0.1528	0.1323	0.1183	0.1080	0.1000	0.0935	0.0882	0.0837
	T.I.	0.4697	0.3180	0.2557	0.2197	0.1955	0.1779	0.1644	0.1535	0.1445	0.1369
6	$\pm 2\sigma$	0.2643	0.1869	0.1526	0.1322	0.1182	0.1079	0.0999	0.0934	0.0881	0.0836
	T.I.	0.4626	0.3152	0.2541	0.2186	0.1947	0.1773	0.1638	0.1530	0.1441	0.1366
7	$\pm 2\sigma$	0.2641	0.1868	0.1525	0.1321	0.1181	0.1078	0.0998	0.0934	0.0880	0.0835
	T.I.	0.4576	0.3133	0.2529	0.2178	0.1941	0.1768	0.1634	0.1527	0.1438	0.1363
8	$\pm 2\sigma$	0.2640	0.1867	0.1524	0.1320	0.1181	0.1078	0.0998	0.0933	0.0880	0.0835
	T.I.	0.4537	0.3118	0.2521	0.2172	0.1937	0.1764	0.1631	0.1524	0.1436	0.1361
9	$\pm 2\sigma$	0.2639	0.1866	0.1523	0.1319	0.1180	0.1077	0.0997	0.0933	0.0880	0.0834
	T.I.	0.4507	0.3106	0.2514	0.2167	0.1933	0.1762	0.1629	0.1522	0.1434	0.1360
10	$\pm 2\sigma$	0.2638	0.1865	0.1523	0.1319	0.1180	0.1077	0.0997	0.0933	0.0879	0.0834
	T.I.	0.4483	0.3097	0.2509	0.2164	0.1931	0.1759	0.1627	0.1521	0.1433	0.1358



Replication Strategy for the Reportable Value

Fusion QbD reports the Components of Variation and the Corresponding % Contributions to Total Analytical Error.

Between Variables Components of Variation

Variable Name	Variance	Standard Deviation	Degrees of Freedom	t-table Value	(+/-) 95% Confidence Limits	Error Contribution (%)
Sample Preparation	0.008	0.092	4	2.7764	0.254	94.17
Injection	0.001	0.023	20	2.0860	0.046	5.83

Overall Error in a Single Determination

Statistic	Value
Mean	100.051
Variance	0.009
Standard Deviation	0.094
% RSD	0.094

Fusion QbD also reports the TOST Results (Traditional Precision Only) and the USP <1210> Interval Results (Combined Precision + Bias).



TOST Analysis Results Summary

Statistic	Value	Pass/Fail
TAE Width (2σ) - Target	±0.300	
Computed TAE Width (2σ)	±0.156	Pass
FPT	<0.0001	
Cp	12.2271	
Variance	0.003	
Standard Deviation	0.055	
% RSD	0.05	
% CV	0.05	

Tolerance Interval Analysis Results

Interval Setting	Value	Number of Preparations	Number of Injections per Preparation
Target	100.000	3	1
Acceptance Limits	±2.000		
Desired Probability %	95.00		
Tolerance Alpha %	5.00		
Grand Mean	100.051		
Computed Tolerance Interval	±0.293	Pass	
Required Guard Band Width	±0.300		

The computed Tolerance Interval falls within the defined Total Analytical Error Limits.



Critical QbD Capability

FMV

Supports All Install Environments (Citrix Ready Certified)



Full 21 CFR Part 11 Compliance Support



Complete Method Validation Experiment Suite



Simple Experiment Workflows




Full LC Experiment Automation



USP 1210> Tolerance and Prediction Interval Metrics





- Replication Strategy and Total Analytical Error
- **Accuracy and Repeatability**
- Analytical Method Transfer




3. ACCURACY AND PRECISION

3.2 Combined Validation of Accuracy and Precision

The illustration below shows that the method will pass System Suitability performance for the Critical Quality Attribute (CQA) being tested SST when Accuracy (β – bias estimate) and Precision (σ – variation estimate) are assessed independently (= High Risk Approach).






3. ACCURACY AND PRECISION

3.2 Combined Validation of Accuracy and Precision

However, as the illustration below shows – the method does not have acceptable System Suitability performance for the Critical Quality Attribute (CQA) being tested when both Accuracy (β – bias estimation) and Precision (σ – variation estimation) **are assessed together (= Low Risk Approach).**



Define your Acceptance Limits:

- Compound-based – USP <1210>
- Computed Results
- Source Data

Small Molecule Data Analysis

Accuracy | Linearity | Repeatability

Select Response for Analysis
Amount

API

Perform Data Analysis

Response Treatment
 % Recovered (Relative) Difference from Mean (Absolute)

Compound-based Acceptance Criteria
 Tolerance / Prediction Interval
 Interval Type
 Tolerance Prediction

Name	Value	Unit
[Specification Limits] <=	0.20	mg
Desired Probability	95.00	%
Tolerance Alpha	5.00	%

Level-based Acceptance Criteria

Computed Results

Target Level	Accuracy [Bias (%) <=]
1.000	5.000
2.000	5.000
4.000	5.000
5.000	5.000
6.000	5.000

Source Data

Set Limit 2.000 %

Mean Value	Individual Result LSL (mg)	Individual Result USL (mg)
1.055	1.034	1.076
2.084	2.043	2.126
4.105	4.023	4.187
5.085	4.984	5.187
5.964	5.845	6.084

The settings are valid.


Automated Reporting – all Results and Graphs for Accuracy, Linearity, Repeatability, and USP <1210> Intervals.

Accuracy Results

Target API (mg)	Mean Observed API - Amount (mg)	Standard Deviation	Lower 95% Confidence Limit	Upper 95% Confidence Limit	RSD (%)	Mean % Bias	Accuracy [% Bias] <=	% Bias Pass/Fail
1.000	1.055	0.031	0.921	1.190	2.96	4.654	5.000	Pass
2.000	2.084	0.032	1.948	2.221	1.52	4.412	5.000	Pass
4.000	4.105	0.012	4.055	4.155	0.28	2.659	5.000	Pass
5.000	5.085	0.021	4.995	5.176	0.41	1.666	5.000	Pass
6.000	5.964	0.039	5.796	6.133	0.66	-0.616	5.000	Pass

Tolerance Interval

Name	Value
Desired Probability %	95.00
Tolerance Alpha %	5.00
Target	0.00
Mean (Pooled)	0.058
Specification Limits (mg)	-0.20 <= Target <= 0.20
Computed Interval (mg)	-0.04 <= Mean <= 0.16
Result	Pass



General Regression Statistics

Regression Statistic Name	Statistic Value
R Square	0.9999
Adj. R Square	0.9999
Residual MSE	682,072,000
Standard Error (±)	26,117
+/- 95% C.I.	56,421
Observations	15

General Validation Acceptance Criteria

Regression Statistic Name	Statistic Value	Validation	Pass / Fail
R	1.0000	0.9998	Pass
Intercept % Bias - Data Based	-0.17	2.00	Pass

Regression Coefficients



Variable Name	Coefficient Value	Coefficient Standard Error	Coefficient t Statistic	P-Value	Lower 95% Confidence Limit	Upper 95% Confidence Limit
Intercept	-11,549	14,735	< 0.0001	0.4472	-43,382	20,283
API	1,715,593	3,638	471.5873	< 0.0001	1,707,734	1,723,452

Natural Variable Model

API - Area (pred) = -11,549 + (1,715,593 x API - Weight Amount)

Range

1.003 (mg) <= API - Weight Amount <= 6.004 (mg)



Critical QbD Capability

FMV

Supports All Install Environments (Citrix Ready Certified)



Full 21 CFR Part 11 Compliance Support



Complete Method Validation Experiment Suite



Simple Experiment Workflows




Full LC Experiment Automation



USP 1210> Tolerance and Prediction Interval Metrics



- Replication Strategy and Total Analytical Error
- Accuracy and Repeatability
- **Analytical Method Transfer**



Comparative Testing

Comparative testing requires the analysis of a predetermined number of samples of the same lot by both the sending and the receiving units. Other approaches may be valid, e.g., if the receiving unit meets a predetermined acceptance criterion for the recovery of an impurity in a spiked product. Such analysis is based on a preapproved transfer protocol that stipulates the details of the procedure, the samples that will be used, **and the predetermined acceptance criteria, including acceptable variability**. Meeting the predetermined acceptance criteria is necessary to assure that the receiving unit is qualified to run the procedure.

Analytical Method Transfer Example

Transferring Lab



Fusion QbD
Sequence
Execution

Chromatography
Data Software

ALR
Design




Chromatogram
Results Data

Fusion QbD
Sequence
Execution

Receiving Lab



1. Fusion QbD – Exports experiment to the CDS as Ready-to-Run sequence, methods, standards
2. Sequence is run at both labs.
3. Fusion QbD – Imports results for instant and complete analysis and reporting.



Accuracy
Linearity
Repeatability
Tolerance Interval
Pass/Fail Results

Key Benefits of FMV

1. Consistency – Workflow and Reporting.

Work is standardized – done the same way every time. Reporting is standardized, complete, easy to communicate.

2. Simplicity

Tremendous ease of use. Very brief learning curve. Clearly defined templatable workflows with built-in workflow management.

3. Speed (Productivity)

Automation and simplified workflows dramatically increase productivity. Review process is minimized and simplified.

4. Regulatory Alignment and Completeness

All required validation experiment types are supported. Reporting meets regulatory requirements. Reports can be attached to Project specific narrative documents.

Key Benefits of FMV

5. Platform Independence

Support for Empower, ChemStation, and Chromeleon means that the standardized workflows and reporting can be easily extended to users of other platforms at other sites or other companies (e.g. CMOs).

6. Customer Support

Our support is top-rated worldwide. S-Matrix and our local distributors have a multi-year history of proven ability to meet all our customer's support needs.

End of Presentation

Analytical Procedure Lifecycle Management Workflow

