

***Mapping Key Elements in the
ICH Q14 and USP <1220> Guidances
to an Enhanced Workflow for
Analytical Procedure Development***

**Richard Verseput
S-Matrix Corporation
www.smatrix.com**



INTERNATIONAL COUNCIL FOR HARMONISATION OF TECHNICAL
REQUIREMENTS FOR PHARMACEUTICALS FOR HUMAN USE

ICH HARMONISED GUIDELINE

ANALYTICAL PROCEDURE DEVELOPMENT Q14

Final Version
Adopted on 1 November 2023

This Guideline has been developed by the appropriate ICH Expert Working Group and has been subject to consultation by the regulatory parties, in accordance with the ICH Process. At Step 4 of the Process the final draft is recommended for adoption to the regulatory bodies of ICH regions.



INTERNATIONAL COUNCIL FOR HARMONISATION OF TECHNICAL
REQUIREMENTS FOR PHARMACEUTICALS FOR HUMAN USE

ICH HARMONISED GUIDELINE

VALIDATION OF ANALYTICAL PROCEDURES Q2(R2)

Final Version
Adopted on 1 November 2023

This Guideline has been developed by the appropriate ICH Expert Working Group and has been subject to consultation by the regulatory parties, in accordance with the ICH Process. At Step 4 of the Process the final draft is recommended for adoption to the regulatory bodies of ICH regions.

Printed on: Thu Mar 07 2024, 04:42:23 PM(EST) Status: Currently Official on 07-Mar-2024 DocId: GUID-35D7E47E-65E5-4567-B4CC-4D96FA230821_2_en-US
Printed by: George Cooney Official Date: Official as of 01-May-2022 Document Type: GENERAL CHAPTER @2024 USPC
Do Not Distribute DOI Ref: 456ba DOI: https://doi.org/10.31003/USP/NF_M10975_Q2_01 1

Add the following:

▲(1220) ANALYTICAL PROCEDURE LIFE CYCLE

INTRODUCTION

This general chapter holistically considers the validation activities that take place across the entire life cycle of an analytical procedure and provides a framework for the implementation of the life cycle approach.

The analytical procedure life cycle approach described here is consistent with the quality by design concepts described in International Council for Harmonisation (ICH) guidelines. The procedure life cycle approach emphasizes the importance of sound scientific approaches and quality risk management for the development, control, establishment, and use of analytical procedures. Total error is used in this chapter; however, measurement uncertainty can also be used.

The procedure life cycle approach is applicable to all types of analytical procedures, and the extent of effort should be consistent with the complexity of the procedure and the criticality of the quality attribute to be measured. The life cycle approach can be considered optional, and any of the elements can be applied on the basis of how the procedure is used. Elements of the life cycle approach can also be applied retrospectively if deemed useful or in early stages of development with the appropriate modifications.

Elements of life cycle management of analytical procedures are also discussed in *Analytical Procedures and Methods Validation for Drugs and Biologics* (Guidance for Industry, FDA 2015).

Validation of an analytical procedure is the process by which it is established, through laboratory studies, that the performance of the procedure meets the requirements for the intended analytical applications. Validation, or demonstration that a procedure is suitable for the intended purpose, takes place during the entire procedure life cycle, beginning during the initial procedure design activities and extending through routine use. These activities include the formal procedure validation, verification, and transfer of procedures, as well as establishing and assuring adherence to an appropriate set of procedure controls and system suitability requirements.

The procedure life cycle is comprised of the analytical target profile (ATP) and three stages, which are introduced below and shown in Figure 1.

The ATP defines the criteria for the procedure performance characteristics that are linked to the intended analytical application and the quality attribute to be measured. It applies to all stages of the procedure life cycle. For quantitative procedures, the ATP describes the required quality of the reportable value since the reportable value generated using a qualified analytical procedure provides the basis for key decisions regarding compliance of a test article with regulatory, compendial, and manufacturing limits. The acceptable level of risk of making an incorrect decision can also be considered when establishing ATP criteria.

Stage 1: Procedure design encompasses procedure development, which consists of the analytical technology and sample preparation. It includes understanding gained through knowledge gathering, systematic procedure development experiments, and risk assessments and associated lab experiments. The output of Stage 1 includes:

1. A set of procedure conditions that minimizes procedure bias to a suitable level, can provide acceptable precision, and can meet the ATP criteria
2. An understanding of the effect of procedure parameters (e.g., temperature, wavelength, flow rate, etc.) on procedure performance
3. Optimization of performance characteristics of the analytical procedure such as accuracy, precision, the appropriateness of any calibration model, specificity and limit of quantitation (as far as applicable); this includes a preliminary replication strategy for samples and standards
4. An initial analytical control strategy (ACS), which is a set of controls (system suitability tests [SSTs] and other procedure-specific controls) needed to ensure proper performance

Stage 2: Procedure performance qualification consists of studies designed to demonstrate that the procedure is suitable for its intended purpose. This involves confirmation that the reportable values generated by application of the analytical procedure meet the ATP criteria as well as confirmation of procedure performance characteristics through the traditional validation, verification, or transfer studies. Data generated during Stage 1 can be used if available and suitable. At the end of Stage 2, the replication strategy and the performance of the procedure is confirmed to meet the ATP and other criteria.


Stage 3: Ongoing procedure performance verification involves monitoring the analytical procedure during routine use and confirming that the performance continues to meet ATP criteria. Monitoring ensures that the performance of the procedure is maintained at an acceptable level over the procedure lifetime. It can also provide an early indication of potential performance issues or adverse trends and aid in identifying required changes for the analytical procedure. Confirming procedure performance after changes ensures that the modified procedure will produce reportable values that meet the criteria defined in the ATP.

More details about the procedure life cycle are described in the subsequent sections.

This presentation will discuss the following key elements within the USP <1210>, USP <1220>, and ICH Q14 guidances which provide a new framework and workflow for analytical procedure development within the context of developing a robust LC method :

- 1. Analytical Target Profile (ATP)**
A Negotiated Specification
- 2. Design of Experiments (DOE, DoE)**
Guided by Risk Assessment
- 3. Method Operable Design Region (MODR)**
Quantitative Robustness Integration
- 4. Replication Strategy Optimization**
Meeting the ATP Performance Requirements

ICH Q14 – Analytical Procedure Lifecycle



Analytical Target Profile

Analytical Target Profile (USP <1220>

The ATP is based on the intended use for the procedure and, for quantitative or semi-quantitative procedures, should include upper limits on the precision and accuracy (bias) of the reportable value.

Example 2: The procedure must be able to quantify [analyte] in a range from [A units] to [B units] in the [description of test article] in the presence of [x, y, z] so that the distribution of the **total analytical error** of the reportable value falls within the **total allowable analytical error range** of \pm [C%].

Negotiation




Replication
Strategy



Total Analytical Error (TAE)

USP <1220>

Total analytical error (TAE) represents the overall error in a test result that is attributed to imprecision and inaccuracy; TAE is the combination of both systematic error of the procedure and random measurement error.





Two Critical Considerations:

1. The Negotiated Total Analytical Error (TAE) **Allowance** for the Analytical Method.
2. The Integration of Precision and Bias into a single **Interval Metric – USP <1210>**.

3. ACCURACY AND PRECISION

3.2 Combined Validation of Accuracy and Precision


An underperforming method can pass System Suitability for the Critical Method Attribute being evaluated when Accuracy (β – bias) and Precision (σ – precision) **are assessed separately = High Risk Approach.**



3. ACCURACY AND PRECISION

3.2 Combined Validation of Accuracy and Precision



As the illustration below shows – the method does not have acceptable System Suitability performance for the Critical Quality Attribute (CQA) being tested when both Accuracy (β – bias estimation) and Precision (σ – variation estimation) are assessed together = Low Risk Approach.



TAE and Guard Bands

Guard Bands acknowledge the presence of Bias and Precision Limits and the need to incorporate a characterized TAE into the “Acceptance Zone”.



Total Analytical Error Distribution



TAE and Guard Bands

Acceptance Zone is narrower to incorporate the characterized TAE.

Total Analytical Error Distribution



TAE and Guard Bands

Production: Amount of Precision-to-Tolerance (P/T) Ratio Available for the Analytical Method

- API method has a tolerance range of 4.0% (i.e., 98.0% to 102.0%)
- Analytical method allowance = 30% of the P/T ratio using a 95% confidence interval.


Determining Required Precision (σ_{\max})

Tolerance Width = 4.00 (98.0 – 102.0)

Precision Width = $0.30 \times 4.00 = 1.20$

Split between LAL and UAL = ± 0.60

$\pm 0.60 = \pm 2\sigma$ width for 95% C.I.



$$2\sigma_{\max} = 0.60$$


Selected Analytical Procedure = LC:

Combined Bias and Precision Allowance Becomes the ATP Quantitation

Performance Metric:

- Robust Method Optimization
- Replication Strategy Optimization

ICH Q14 – Analytical Procedure Lifecycle



Risk Assessment



ICH Q14

Risk assessment and prior knowledge should be used to identify analytical procedure parameters, attributes and associated ranges to be investigated experimentally. Categorical variables (e.g., different instruments) can also be considered as part of the experimental design.

USP <1220>

For variables where there may be higher risk, one way to reduce risk is to gain additional knowledge about the influence of those parameters using modeling and/or experimentation.

Sources of Risk for Bias and Variation


Element	Presumed CMPs	CMAs						Category (C, N, X)
		Resolution USP	S/N	Tailing USP	Area % RSD - API	K-Prime - 1st Peak	K-Prime - Last Peak	
Chemistry	Column Type	5	1	1	3	5	5	X-S
	Strong Solvent	5	1	1	3	5	5	X-S
	Aqueous solvent	5	5	5	1	5	5	X-S
	pH	5	5	5	3	5	5	X-S-O
Process	Pump Flow Rate	3	1	5	3	5	5	X-O
	Injection Volume	3	5	3	5	1	1	C
	Oven Temperature	5	1	3	3	5	5	X-O
Gradient Program	Initial Hold Time	1	1	1	1	5	1	C-O
	Gradient Slope	5	1	5	3	5	5	X-S-O
Detection	Wavelength	5	5	1	5	1	1	C
	Sampling Rate	3	5	1	5	1	1	C
	Precision	1	3	1	3	1	1	C

C = Controlled Factor, N = Noise Factor, X = eXperimental Factor (S = Screening, O = Optimization)

Impact Severity

Low = 1
Medium = 3
High = 5

Design of Experiments (DOE, DoE)




ICH Q14

In an enhanced approach, the ranges for the relevant parameters and their interactions can be investigated in multivariate experiments (DoE).

USP <1220>

Experimentation is a direct way of generating data that can be used to assess the impact of procedure parameters on performance, and the use of statistical design of experiments (DOE) is an effective way to do this.



DoE uses Statistical Sampling of All Possible Combinations to Support Accurate Estimation of Study Factor Effects



$$R_S = b_0 + b_1(x1) + b_2(x2) + b_{11}(x1)^2 + b_{22}(x2)^2 + +b_{12}(x1 * x2)$$

DoE (DOE) – A Model Building Methodology

Run	Factor 1	Factor 2	Factor 3	Response	Model	Residual	Std. Error	95% CI	95% PI	Optimal	Quality	Remarks
1	1.00	1.00	1.00	1.00	1.00	1.00	1.00	1.00	1.00	1.00	1.00	
2	1.00	1.00	2.00	2.00	2.00	2.00	2.00	2.00	2.00	2.00	2.00	
3	1.00	1.00	3.00	3.00	3.00	3.00	3.00	3.00	3.00	3.00	3.00	
4	1.00	1.00	4.00	4.00	4.00	4.00	4.00	4.00	4.00	4.00	4.00	
5	1.00	1.00	5.00	5.00	5.00	5.00	5.00	5.00	5.00	5.00	5.00	
6	1.00	2.00	1.00	6.00	6.00	6.00	6.00	6.00	6.00	6.00	6.00	
7	1.00	2.00	2.00	7.00	7.00	7.00	7.00	7.00	7.00	7.00	7.00	
8	1.00	2.00	3.00	8.00	8.00	8.00	8.00	8.00	8.00	8.00	8.00	
9	1.00	2.00	4.00	9.00	9.00	9.00	9.00	9.00	9.00	9.00	9.00	
10	1.00	2.00	5.00	10.00	10.00	10.00	10.00	10.00	10.00	10.00	10.00	
11	1.00	3.00	1.00	11.00	11.00	11.00	11.00	11.00	11.00	11.00	11.00	
12	1.00	3.00	2.00	12.00	12.00	12.00	12.00	12.00	12.00	12.00	12.00	
13	1.00	3.00	3.00	13.00	13.00	13.00	13.00	13.00	13.00	13.00	13.00	
14	1.00	3.00	4.00	14.00	14.00	14.00	14.00	14.00	14.00	14.00	14.00	
15	1.00	3.00	5.00	15.00	15.00	15.00	15.00	15.00	15.00	15.00	15.00	
16	1.00	4.00	1.00	16.00	16.00	16.00	16.00	16.00	16.00	16.00	16.00	
17	1.00	4.00	2.00	17.00	17.00	17.00	17.00	17.00	17.00	17.00	17.00	
18	1.00	4.00	3.00	18.00	18.00	18.00	18.00	18.00	18.00	18.00	18.00	
19	1.00	4.00	4.00	19.00	19.00	19.00	19.00	19.00	19.00	19.00	19.00	
20	1.00	4.00	5.00	20.00	20.00	20.00	20.00	20.00	20.00	20.00	20.00	
21	1.00	5.00	1.00	21.00	21.00	21.00	21.00	21.00	21.00	21.00	21.00	
22	1.00	5.00	2.00	22.00	22.00	22.00	22.00	22.00	22.00	22.00	22.00	
23	1.00	5.00	3.00	23.00	23.00	23.00	23.00	23.00	23.00	23.00	23.00	
24	1.00	5.00	4.00	24.00	24.00	24.00	24.00	24.00	24.00	24.00	24.00	
25	1.00	5.00	5.00	25.00	25.00	25.00	25.00	25.00	25.00	25.00	25.00	



$$\# - R_s \geq 2.00 = 9.3 + 4.2(\text{PFR}) - 5.4(\Delta t_G)^2 + 12.7(\Delta t_G * \text{pH}) + 1.3(\text{pH} * \Delta T) + 1.6[(\Delta T)^2 * \Delta t_G] + \dots$$

Linear Effect Curvature Effect

Interaction Effects

Complex Effect

Parameter Selection – Screening Study

Method Parameter	Study Range
pH	2.70 – 4.90
Gradient Time (min)	10.0 – 25.0
Column Type	BEH C18 BEH Shield RP18 HSS T3 CSH Phenyl-Hexyl

Prior knowledge (from original monograph) incorporated into Selected Column and Chemistry Study Factors and Range.

Fusion QbD Experiment Automation – Export to Empower



Generates QbD-aligned
DOE Experiment

Automatically Builds
Sequence and All
Instrument Methods

Empower CDS

RD2 Optimization in S-Matrix\Internal Development\FMD\RD2 - Optimization - 9_9_0 as System/Administrator - Sample Set Method Editor

File Edit View Help

Apply Table Preferences Sample Set Method

Plate/Well	Inj Vol (uL)	# of Injs	Label	SampleName	Level	Function	Method Set / Report or Export Method	Label Reference	Processing	Run Time (Minutes)	Data Start (Minutes)	Next Inj Delay (Minutes)	MS Tune Method	MS Calibration Method	Column Position	Auto Additions	SampleWeight	Dilution
1						Condition Column	RD2 Optimization 001_031			6.60					Position 5			
2						Equilibrate	RD2 Optimization 001_001			4.00					No Change			
3	1:A,1	2.0	1	Unk-001-001	1	Inject Samples	RD2 Optimization 001_001		Normal	17.60	0.00	3.50					1.00000	1.00000
4						Equilibrate	RD2 Optimization 001_002			4.00					No Change			
5	1:A,1	2.0	1	Unk-001-002	2	Inject Samples	RD2 Optimization 001_002		Normal	9.60	0.00	3.50					1.00000	1.00000
6						Equilibrate	RD2 Optimization 001_003			4.00					No Change			
7	1:A,1	2.0	1	Unk-001-003	3	Inject Samples	RD2 Optimization 001_003		Normal	17.60	0.00	3.50					1.00000	1.00000
8						Equilibrate	RD2 Optimization 001_004			4.00					No Change			
9	1:A,1	2.0	1	Unk-001-004	4	Inject Samples	RD2 Optimization 001_004		Normal	9.60	0.00	3.50					1.00000	1.00000
10						Equilibrate	RD2 Optimization 001_005			4.00					No Change			
11	1:A,1	2.0	1	Unk-001-005	5	Inject Samples	RD2 Optimization 001_005		Normal	9.60	0.00	3.50					1.00000	1.00000
12						Equilibrate	RD2 Optimization 001_006			4.00					No Change			
13	1:A,1	2.0	1	Unk-001-006	6	Inject Samples	RD2 Optimization 001_006		Normal	9.60	0.00	3.50					1.00000	1.00000
14						Equilibrate	RD2 Optimization 001_007			4.00					No Change			
15	1:A,1	2.0	1	Unk-001-007	7	Inject Samples	RD2 Optimization 001_007		Normal	17.60	0.00	3.50					1.00000	1.00000
16						Condition Column	RD2 Optimization 001_032			6.60					Position 5			
17						Equilibrate	RD2 Optimization 001_008			4.00					No Change			
18	1:A,1	2.0	1	Unk-001-008	8	Inject Samples	RD2 Optimization 001_008		Normal	13.60	0.00	3.50					1.00000	1.00000
19						Condition Column	RD2 Optimization 001_033			6.60					Position 5			
20						Equilibrate	RD2 Optimization 001_009			4.00					No Change			
21	1:A,1	2.0	1	Unk-001-009	9	Inject Samples	RD2 Optimization 001_009		Normal	9.60	0.00	3.50					1.00000	1.00000
22						Equilibrate	RD2 Optimization 001_010			4.00					No Change			
23	1:A,1	2.0	1	Unk-001-010	10	Inject Samples	RD2 Optimization 001_010		Normal	9.60	0.00	3.50					1.00000	1.00000
24						Equilibrate	RD2 Optimization 001_011			4.00					No Change			
25	1:A,1	2.0	1	Unk-001-011	11	Inject Samples	RD2 Optimization 001_011		Normal	17.60	0.00	3.50					1.00000	1.00000
26						Equilibrate	RD2 Optimization 001_012			4.00					No Change			
27	1:A,1	2.0	1	Unk-001-012	12	Inject Samples	RD2 Optimization 001_012		Normal	17.60	0.00	3.50					1.00000	1.00000
28						Condition Column	RD2 Optimization 001_034			6.60					Position 5			
29						Equilibrate	RD2 Optimization 001_013			4.00					No Change			
30	1:A,1	2.0	1	Unk-001-013	13	Inject Samples	RD2 Optimization 001_013		Normal	13.60	0.00	3.50					1.00000	1.00000
31						Condition Column	RD2 Optimization 001_035			6.60					Position 5			
32						Equilibrate	RD2 Optimization 001_014			4.00					No Change			
33	1:A,1	2.0	1	Unk-001-014	14	Inject Samples	RD2 Optimization 001_014		Normal	13.60	0.00	3.50					1.00000	1.00000
34						Equilibrate	RD2 Optimization 001_015			4.00					No Change			
35	1:A,1	2.0	1	Unk-001-015	15	Inject Samples	RD2 Optimization 001_015		Normal	13.60	0.00	3.50					1.00000	1.00000
36						Equilibrate	RD2 Optimization 001_016			4.00					No Change			
37	1:A,1	2.0	1	Unk-001-016	16	Inject Samples	RD2 Optimization 001_016		Normal	9.60	0.00	3.50					1.00000	1.00000
38						Equilibrate	RD2 Optimization 001_017			4.00					No Change			

For Help, press F1

Fusion QbD Experiment Automation

– Import from Empower



Fusion QbD®


Automatically Retrieves All Chromatogram Results Data

Automated analysis, graphing, and reporting.


Report formats:

RTF, DOC, HTML, PDF, XLSX, XML

Empower CDS




Plate/Vial	Wt (µL)	# of Inj	Label	Sample Name	Level	Function	Method Set / Report or Export Method	Label Reference	Processing	Run Time (Minutes)	Data Start (Minutes)	Inject Delay (Minutes)	MS Tune Method	MS Calibration Method	Column Position	Auto Addition	Sample Weight	Dilution
1						Condition Column	RD2 Optimization 001_031			6.00					Position 5			
2						Equilibrate	RD2 Optimization 001_031			4.00					No Change			
3	1.A.1	2.0	1	Unk-001-001.1		Inject Samples	RD2 Optimization 001_001		Normal	17.00	0.00	3.50			No Change		1.00000	1.00000
4						Equilibrate	RD2 Optimization 001_002			4.00					No Change			
5	1.A.1	2.0	1	Unk-001-002.2		Inject Samples	RD2 Optimization 001_002		Normal	9.00	0.00	3.50			No Change		1.00000	1.00000
6						Equilibrate	RD2 Optimization 001_003			4.00					No Change			
7	1.A.1	2.0	1	Unk-001-003.3		Inject Samples	RD2 Optimization 001_003		Normal	17.00	0.00	3.50			No Change		1.00000	1.00000
8						Equilibrate	RD2 Optimization 001_004			4.00					No Change			
9	1.A.1	2.0	1	Unk-001-004.4		Inject Samples	RD2 Optimization 001_004		Normal	9.00	0.00	3.50			No Change		1.00000	1.00000
10						Equilibrate	RD2 Optimization 001_005			4.00					No Change			
11	1.A.1	2.0	1	Unk-001-005.5		Inject Samples	RD2 Optimization 001_005		Normal	9.00	0.00	3.50			No Change		1.00000	1.00000
12						Equilibrate	RD2 Optimization 001_006			4.00					No Change			
13	1.A.1	2.0	1	Unk-001-006.6		Inject Samples	RD2 Optimization 001_006		Normal	9.00	0.00	3.50			No Change		1.00000	1.00000
14						Equilibrate	RD2 Optimization 001_007			4.00					No Change			
15	1.A.1	2.0	1	Unk-001-007.7		Inject Samples	RD2 Optimization 001_007		Normal	17.00	0.00	3.50			No Change		1.00000	1.00000
16						Condition Column	RD2 Optimization 001_032			6.00					Position 5			
17						Equilibrate	RD2 Optimization 001_008			4.00					No Change			
18	1.A.1	2.0	1	Unk-001-008.8		Inject Samples	RD2 Optimization 001_008		Normal	12.00	0.00	3.50			No Change		1.00000	1.00000
19						Condition Column	RD2 Optimization 001_033			6.00					Position 5			
20						Equilibrate	RD2 Optimization 001_009			4.00					No Change			
21	1.A.1	2.0	1	Unk-001-009.9		Inject Samples	RD2 Optimization 001_009		Normal	9.00	0.00	3.50			No Change		1.00000	1.00000
22						Equilibrate	RD2 Optimization 001_010			4.00					No Change			
23	1.A.1	2.0	1	Unk-001-010.10		Inject Samples	RD2 Optimization 001_010		Normal	9.00	0.00	3.50			No Change		1.00000	1.00000
24						Equilibrate	RD2 Optimization 001_011			4.00					No Change			
25	1.A.1	2.0	1	Unk-001-011.11		Inject Samples	RD2 Optimization 001_011		Normal	17.00	0.00	3.50			No Change		1.00000	1.00000
26						Equilibrate	RD2 Optimization 001_012			4.00					No Change			
27	1.A.1	2.0	1	Unk-001-012.12		Inject Samples	RD2 Optimization 001_012		Normal	17.00	0.00	3.50			No Change		1.00000	1.00000
28						Condition Column	RD2 Optimization 001_034			6.00					Position 5			
29						Equilibrate	RD2 Optimization 001_013			4.00					No Change			
30	1.A.1	2.0	1	Unk-001-013.13		Inject Samples	RD2 Optimization 001_013		Normal	15.00	0.00	3.50			No Change		1.00000	1.00000
31						Condition Column	RD2 Optimization 001_035			6.00					Position 5			
32						Equilibrate	RD2 Optimization 001_014			4.00					No Change			
33	1.A.1	2.0	1	Unk-001-014.14		Inject Samples	RD2 Optimization 001_014		Normal	13.00	0.00	3.50			No Change		1.00000	1.00000
34						Equilibrate	RD2 Optimization 001_015			4.00					No Change			
35	1.A.1	2.0	1	Unk-001-015.15		Inject Samples	RD2 Optimization 001_015		Normal	15.00	0.00	3.50			No Change		1.00000	1.00000
36						Equilibrate	RD2 Optimization 001_016			4.00					No Change			
37	1.A.1	2.0	1	Unk-001-016.16		Inject Samples	RD2 Optimization 001_016		Normal	9.00	0.00	3.50			No Change		1.00000	1.00000
38						Equilibrate	RD2 Optimization 001_017			4.00					No Change			




Screening Study – Simple Analysis



Quantitative Knowledge



Required Performance Threshold



Region of Acceptable Performance
– One Response



Shaded region = methods that do NOT meet performance requirements.

Screening Study – Simple Analysis

Reports
Optimization Study Ranges: APR_2
View as Report

Graph Settings

Name	Units	Lower Bound	Upper Bound	Pointer Coordinate
X Gradient Time	min	20.0	30.0	25.0
Y pH	°	3.400	4.400	4.000

Column Type: CSH Phenyl Hexyl

Optimization – CSH Phenyl-Hexyl Column


Rectangle Identifies the pH and t_G Study Ranges to use in an Optimization Experiment with the Phenyl-Hexyl Column

Gradient Time	20.1	29.9	25.0	25.0
pH	3.773	4.224	3.998	4.000

Verification Runs: None
 Include Verification Runs in Report
 Show Verification Run Labels
 Include Prediction Chromatograms in Report

Run ID	Gradient Time	pH	Column Type

Graph: Fusion QbD Graph



Simple Color-coded Responses. Shaded Regions Correspond to Under-performing Methods

Overlay: Rs-Map


Response Settings

Name	Units	Goal	Lower Bound	Upper Bound	Color	Crosshair Prediction	Contour Label
No. of Peaks	*	Target	7.0	9.0	Gray	7.72	✓
No. of Peaks >= 1.50 - USPResolution		Target	6.0	8.0	Red	6.23	✓
No. of Peaks >= 2.00 - USPResolution		Target	6.0	8.0	Blue	6.06	✓
No. of Peaks <= 1.50 - USPTailing		Target	7.0	9.0	Green	7.73	✓
First Peak - RetentionTime	min	Maximize	0.90		Orange	1.048	✓


Method Parameter	Study Range
Pump Flow Rate (mL/min)	0.35 – 0.55
Column Oven Temperature (°C)	40.0 – 50.0
Gradient Time (min)*	8.0 – 16.0
pH	3.60 – 4.20
Column Type	CSH Phenyl-Hexyl

Light green background color indicates result obtained from screening study.


* – t_G range adjusted to maintain slope range with new endpoint of 50% ACN.



UV Chromatogram



MS Chromatogram



Peak Table - PDA Ch1 225nm@4.8nm - Run No. 11

Name	RT (min)	Base Peak (m/z)	Height (uV)	Area
1 B	1.916	262.2	524,998	1,508,961
2 G	5.336	193.1	113,225	3,8630
3 H*	5.720	144.2	403,264	20,577,007
4 F	6.400	336.8	185,768	4,7804
5 A	6.889	274.2	260,696	6,11066
6 API	7.139	288.2	774,666	2,377,011
7 D-Deg	7.139	170.503	170,503	
8 E	7.747	304.1	522,540	1,276,777
9 D	9.748	360.2	212,243	59,618

Global Tracking Method (GTM): Generated from run 15

Mass UV Filters Auto-name Peaks in GTM

Display Intensity Columns

Run #	Component Name	RT	Area	Expected Mass 1	Expected Mass 2	Expected Mass 3	Expected Mass 4	Expected Mass 5
1	15 B	1.037	973,202.2	262.3	263.2	279.1	160.3	126.1
2	15 G	3.061	110,123.2	193.1	215.1	132.1	279.1	194.1
3	15* H*	3.958	17,435,187.9	144.2	190.2	145.3	126.0	191
4	15 F	4.656	296,880.8	336.8	672.4	673.5	279.3	
5	15 A	4.926	397,548.5	274.2	275.2	279.0	296.2	126
6	15 API	5.122	1,247,108.7	288.1	289.1	126.0	160.3	310
7	15 D-Deg	5.844	200,691.0	300.2	341.0	126.0	279.1	303
8	15 E	5.633	809,240.9	304.1	305.2	326.1	607.5	279
9	15 D	7.171	332,729.9	360.2	361.2	279.1	126.0	160

UV and MS Spectra Analysis Dialog

Extracted Spectra: Row 7 (7)

Detected Mass (Da)	Leading (Intensity)	Apex (Intensity)	Trailing (Intensity)
1 661	4,201,244	13,400,470	598,699
2 663	2,442,044	10,676,912	493,867
3 662	1,339,995	3,645,166	0
4 665	856,023	2,571,824	47,855
5 683	270,540	825,603	0
6 684	289,782	365,052	3,800
7 460	95,431	176,563	11,209
8 685	56,250	170,395	31,053

Detected Wavelength	Leading (AU)	Apex (AU)	Trailing (AU)
1 215.8	0.000	1.239	0.000
2 251.3	0.038	0.322	0.000
3 259.1	0.000	0.000	0.002
4 253.7	0.000	0.000	0.002
5 189.7	0.000	0.202	0.000
6 282.6	0.000	0.140	0.000
7 282.0	0.016	0.000	0.000

Intensity Threshold Options: Show All Above Threshold: 0 Mass Precision 0 Intensity Precision 0 Refresh

Automatic Peak Deconvolution

Virtual Integration of Non-absorbing Peaks


Select Mode

Data Review Peak Tracking

Commands


Create Tracking Method Track Peaks... Apply Tracking Changes Close

Traditional Resolution Map – 2D



**Composite UV Chromatogram
Includes Virtual MS Peak**

2D Contour Graph



Resolution Map – Rotatable 3D

Reports

Robust Design Space 1 | APR_2 | Update Rs-Map

View as Report


Graph Settings

Name	Units	Lower Bound	Upper Bound	Pointer Coordinate
X pH	*	3.600	4.200	3.800
Y Oven Temperature	°C	40.0	50.0	45.0

Pump Flow Rate: 0.400
Gradient Time: 15.0

Simulation Chromatogram

APR_2 - Pointer Coordinate | Fusion QbD Predicted Chromatogram




UV

Minutes

3D Response Surface Graph

Fusion QbD Graph



Legend

4.17
3.95
3.73
3.51
3.30
3.08
2.86
2.64
2.42
2.20
1.98
1.76
1.54
1.32
1.10
0.88
0.66
0.44
0.22
0.00

Overlay Rs-Map

Resolution Map Settings

Name	Pointer Predictions
2 API	4.118
3 D-Deg	5.670
4 E	6.218
5 A	7.784
6 F	13.326
7 H*	17.583
8 G	32.760
9 D	33.185

View as 3D | Reset

Graph Density

Low (Fast)
 Medium (Medium)
 High (Slow)

Color Mode



Color
 Greyscale
 B & W

3D Rs-Map Commands

Print... | Copy

Mean Performance:

- Resolution
- K-Prime
- Tailing
- Area %RSD
- Etc.



Integrated Robustness Simulation



ICH Q14

Data gained during the development studies (e.g., robustness data from a design of experiments (DoE) study) could be used as part of the validation data for the related analytical procedure performance characteristics and studies do not necessarily need to be repeated.

USP <1220>

In some cases, it is helpful to demonstrate robustness of the procedure by developing models that describe the effect of parameters on the performance of the procedure, ... This knowledge also enables the determination of robust operation regions for procedure parameters and, if desired, a method operable design region (MODR).

Wizard Page 1 – Define Maximum Expected Variation in Study Parameters

Variable Settings			
Enabled	Experiment Variable	Units	Maximum Expected Variation ($\pm 3\sigma$ Value)
<input checked="" type="checkbox"/>	Pump Flow Rate	mL/min	0.020
<input checked="" type="checkbox"/>	Oven Temperature	°C	3.0
<input checked="" type="checkbox"/>	pH	*	0.15
<input checked="" type="checkbox"/>	Mobile Phase Composition (MPC)*	%	2.0

* - MPC variation is composition (blend) variation due to pump precision limits. A commonly used $\pm 3\sigma$ value = $\pm 2.0\%$.
 The value you enter will be applied to all Gradient Slope factors (e.g., Time, Slope, and Ramp Steps) in the experiment design.

Go beyond development LC system

expected variation in QC lab across LC systems during routine use.




Wizard Page 2 – Define Failure Mode and Spec Limits

Response Settings

Additional Error
C.I. for Simulation 2 Sigma ▾

Enabled	Response	Robustness Index	Specification Limit Delta (\pm)	LSL	USL	Target	Additional Error	Additional Error Amount ($\pm 1\sigma$ Value)
<input checked="" type="checkbox"/>	Rs-Map Response	Cpk ▾		1.500				
<input checked="" type="checkbox"/>	B - RetentionTime	Cpk ▾		1.00			<input type="checkbox"/>	
<input checked="" type="checkbox"/>	API - USPTailing	Cpk ▾			1.50		<input type="checkbox"/>	
<input checked="" type="checkbox"/>	A - ResolutionW50	Cpk ▾		1.500				
<input checked="" type="checkbox"/>	API - ResolutionW50	Cpk ▾		1.500				
<input checked="" type="checkbox"/>	D-Deg - ResolutionW50	Cpk ▾		1.500				
<input checked="" type="checkbox"/>	E - ResolutionW50	Cpk ▾		1.500				

Monte Carlo Robustness Simulation



Multi-Response Overlay View

Reports

Robust Design Space 1 | APR 2 | Update Rs-Map

View as Report


Graph Settings

Name	Units	Lower Bound	Upper Bound	Pointer Coordinate
X pH	*	3.60	4.20	3.80
Y Oven Temperature	°C	40.0	50.0	45.0

Pump Flow Rate: 0.400
Gradient Time: 15.0


Simulation Chromatogram

Fusion QbD Predicted Chromatogram



Validation Status: Your settings are valid.

Graph



Overlay | Rs-Map


Response Settings

Add Named Peak Rs Responses...

Name	Units	Goal	Lower Bound	Upper Bound	Crosshair Prediction	Contour Label	Color
A - ResolutionW50	*	Maximize	2.000		7.784		Red
API - ResolutionW50	*	Maximize	2.000		4.118		Blue
D-Deg - ResolutionW50	*	Maximize	2.000		5.670		Green
E - ResolutionW50	*	Maximize	2.000		6.218		Orange
B - RetentionTime		Maximize	1.00		1.341		Gray
API - USPTailing		Minimize		1.50	1.345		Purple
B - RetentionTime - Cpk	*	Maximize	1.330		2.2410		Gray
API - USPTailing - Cpk	*	Maximize	1.330		4.8516		Purple

4-Factor Method Operable Design Region (MODR).

- MODR (unshaded region) – methods are robust for all CQAs.
- Rectangle – independently adjustable ranges within which permanent post-approval changes can be made while maintaining robust performance for all CQAs.



Replication Strategy Optimization



ICH Q14

Reportable Result: the result as generated by the analytical procedure after calculation or processing and applying the described sample replication. (ICH Q2)

ICH Q2(R2)

The experimental design of the validation study should reflect the number of replicates used in routine analysis to generate a reportable result.

USP <1220>

Stage 1:

Optimization of performance characteristics of the analytical procedure such as accuracy, precision, ...; this includes a preliminary replication strategy for samples and standards.

Method Development - Untitled1

File Edit Activity Tools Window Help

Select Autosampler Tray Update Setup Data Generate Design ?

Design of Experiments

- Create a Design
- Design Reports

Data Entry / Analysis

- Data Entry
- Data Analysis

Reporting Toolkit

- Fusion Reporter
- Audit Log Reporter

Project Name: Project 1 Experiment Name: Experiment 1 Instrument Name: Fusion QbD H_Class Experiment Phase: Method Development Experiment Type: Replication Strategy Separation Mode: Reversed Phase (RPC)

Experiment Setup

Global Sample Settings

Obtain all injection repeats from the same vial

Name: Preparation replicates per sample No. of Levels: 5

	Level setting
Level 1	P-1
Level 2	P-2
Level 3	P-3
Level 4	P-4
Level 5	P-5

Name: Injections per preparation replicate No. of Levels: 5

	Level setting
Level 1	I-1
Level 2	I-2
Level 3	I-3
Level 4	I-4
Level 5	I-5

TAE and Guard Bands

Production: Amount of Precision-to-Tolerance (P/T) Ratio Available for the Analytical Method

- API method has a tolerance range of 4.0% (i.e., 98.0% to 102.0%)
- Analytical method allowance = 30% of the P/T ratio using a 95% confidence interval.


Determining Required Precision (σ_{\max})

Tolerance Width = 4.00 (98.0 – 102.0)

Precision Width = $0.30 \times 4.00 = 1.20$

Split between LAL and UAL = ± 0.60

$\pm 0.60 = \pm 2\sigma$ width for 95% C.I.



Replication Strategy for the Reportable Value

Between Variables Components of Variation

Variable Name	Variance	Standard Deviation	Degrees of Freedom	t-table Value	(+/-) 95% Confidence Limits	Error Contribution (%)
Sample Preparation	0.065	0.256	4	2.7764	0.71	95.27
Injection	0.003	0.057	20	2.0860	0.11	4.73

Overall Error in a Single Determination

Statistic	Value
Mean	100.142
Variance	0.069
Standard Deviation	0.262
% RSD	0.262


No. of Injections	No. of Preparations										
	1	2	3	4	5	6	7	8	9	10	
1	$\pm 2\sigma$	0.7517	0.5311	0.4340	0.3759	0.3362	0.3069	0.2841	0.2658	0.2506	0.2377
	T.I.	1.7228	1.0551	0.8121	0.6810	0.5968	0.5372	0.4922	0.4568	0.4280	0.4040
2	$\pm 2\sigma$	0.7428	0.5252	0.4288	0.3714	0.3322	0.3032	0.2807	0.2626	0.2476	0.2349
	T.I.	1.4742	0.9516	0.7506	0.6383	0.5646	0.5115	0.4709	0.4387	0.4122	0.3900
3	$\pm 2\sigma$	0.7398	0.5231	0.4271	0.3699	0.3308	0.3020	0.2796	0.2615	0.2466	0.2339
	T.I.	1.3843	0.9156	0.7296	0.6239	0.5537	0.5028	0.4638	0.4326	0.4069	0.3854
4	$\pm 2\sigma$	0.7383	0.5220	0.4262	0.3691	0.3302	0.3014	0.2790	0.2610	0.2461	0.2335
	T.I.	1.3376	0.8973	0.7189	0.6166	0.5482	0.4985	0.4602	0.4295	0.4043	0.3830
5	$\pm 2\sigma$	0.7374	0.5214	0.4257	0.3687	0.3298	0.3010	0.2787	0.2607	0.2458	0.2332
	T.I.	1.3089	0.8862	0.7125	0.6122	0.5450	0.4959	0.4580	0.4277	0.4027	0.3816
6	$\pm 2\sigma$	0.7368	0.5210	0.4254	0.3684	0.3295	0.3008	0.2785	0.2605	0.2456	0.2330
	T.I.	1.2896	0.8787	0.7082	0.6093	0.5428	0.4941	0.4566	0.4265	0.4016	0.3807
7	$\pm 2\sigma$	0.7363	0.5207	0.4251	0.3682	0.3293	0.3006	0.2783	0.2603	0.2454	0.2328
	T.I.	1.2756	0.8733	0.7051	0.6072	0.5412	0.4929	0.4556	0.4256	0.4009	0.3800
8	$\pm 2\sigma$	0.7360	0.5204	0.4249	0.3680	0.3291	0.3005	0.2782	0.2602	0.2453	0.2327
	T.I.	1.2650	0.8693	0.7028	0.6056	0.5400	0.4920	0.4548	0.4250	0.4003	0.3795
9	$\pm 2\sigma$	0.7357	0.5202	0.4248	0.3679	0.3290	0.3004	0.2781	0.2601	0.2452	0.2327
	T.I.	1.2568	0.8662	0.7010	0.6044	0.5391	0.4912	0.4542	0.4244	0.3999	0.3791
10	$\pm 2\sigma$	0.7355	0.5201	0.4247	0.3678	0.3289	0.3003	0.2780	0.2601	0.2452	0.2326
	T.I.	1.2501	0.8636	0.6995	0.6034	0.5384	0.4906	0.4537	0.4240	0.3995	0.3788

TOST Analysis Results Summary


Statistic	Value	Pass/Fail
TAE Width (2σ) - Target	± 0.600	
Computed TAE Width (2σ)	± 0.434	Pass
FPT	<0.0001	
Cp	4.4075	
Variance	0.023	
Standard Deviation	0.151	
% RSD	0.15	
% CV	0.15	

Tolerance Interval Analysis Results

Interval Setting	Value	Number of Preparations	Number of Injections per Preparation
Target	100.000	3	1
Acceptance Limits	± 2.000		
Desired Probability %	95.00		
Tolerance Alpha %	5.00		
Grand Mean	100.142		
Computed Tolerance Interval	± 0.812		Fail
Required Guard Band Width	± 0.600		



Sample Prep Method Optimization



ICH Q14

A sample and/or sample preparation is considered suitable if the measurement response of the sample satisfies pre-defined acceptance criteria for the analytical procedure attributes that have been developed for the validated analytical procedure.

USP <1220>


Stage 1: Procedure design encompasses procedure development, which consists of the analytical technology and sample preparation.



Sample Preparation Method Optimization

Name	Units	Type	Level Settings	
Buffer pH	°	Discrete Numeric	Level 1	8.00
		No. of Levels		3
		Level 2	8.50	
		Level 3	9.00	
State: <input checked="" type="radio"/> Variable <input type="radio"/> Constant				
Name	Units	Type	Lower Bound	Upper Bound
Organic Level	%	Continuous	20	50
State: <input checked="" type="radio"/> Variable <input type="radio"/> Constant				
Name	Units	Type	Lower Bound	Upper Bound
Sorption Time	min	Continuous	0	30
State: <input checked="" type="radio"/> Variable <input type="radio"/> Constant				
Name	Units	Type	Lower Bound	Upper Bound
Shaker Speed	rpm	Continuous	50	250
State: <input checked="" type="radio"/> Variable <input type="radio"/> Constant				
Name	Units	Type	Lower Bound	Upper Bound
Shaker Time	min	Continuous	20	120
State: <input checked="" type="radio"/> Variable <input type="radio"/> Constant				

Optimization reduces the amount of the TAE contributed by Sample Preparation.



Replication Strategy for the Reportable Value

Between Variables Components of Variation

Variable Name	Variance	Standard Deviation	Degrees of Freedom	t-table Value	(+/-) 95% Confidence Limits	Error Contribution (%)
Sample Preparation	0.029	0.170	4	2.7764	0.471	96.09
Injection	0.001	0.034	20	2.0860	0.071	3.91

Overall Error in a Single Determination

Statistic	Value
Mean	100.091
Variance	0.030
Standard Deviation	0.173
% RSD	0.173


TOST Analysis Results Summary

Statistic	Value	Pass/Fail
TAE Width (2σ) - Target	±0.600	
Computed TAE Width (2σ)	±0.287	Pass
FPT	<0.0001	
Cp	6.6753	
Variance	0.010	
Standard Deviation	0.100	
% RSD	0.10	
% CV	0.10	


Tolerance Interval Analysis Results


Interval Setting	Value	Number of Preparations	Number of Injections per Preparation
Target	100.000	3	1
Acceptance Limits	±2.000		
Desired Probability %	95.00		
Tolerance Alpha %	5.00		
Grand Mean	100.091		
Computed Tolerance Interval	±0.536	Pass	
Required Guard Band Width	±0.600		

No. of Injections		No. of Preparations									
		1	2	3	4	5	6	7	8	9	10
1	±2σ	0.4963	0.3511	0.2866	0.2482	0.2220	0.2026	0.1876	0.1755	0.1654	0.1570
	T.I.	1.1375	0.6961	0.5362	0.4496	0.3940	0.3547	0.3250	0.3016	0.2826	0.2668
2	±2σ	0.4915	0.3475	0.2837	0.2457	0.2198	0.2006	0.1858	0.1738	0.1638	0.1554
	T.I.	0.9754	0.6296	0.4967	0.4224	0.3736	0.3384	0.3116	0.2902	0.2727	0.2581
3	±2σ	0.4898	0.3464	0.2828	0.2449	0.2191	0.2000	0.1851	0.1732	0.1633	0.1549
	T.I.	0.9166	0.6063	0.4831	0.4131	0.3666	0.3329	0.3071	0.2864	0.2694	0.2552
4	±2σ	0.4890	0.3458	0.2823	0.2445	0.2187	0.1996	0.1848	0.1729	0.1630	0.1546
	T.I.	0.8860	0.5943	0.4762	0.4084	0.3631	0.3302	0.3048	0.2845	0.2678	0.2537
5	±2σ	0.4885	0.3454	0.2820	0.2443	0.2185	0.1994	0.1846	0.1727	0.1628	0.1545
	T.I.	0.8672	0.5871	0.4720	0.4056	0.3610	0.3285	0.3035	0.2834	0.2668	0.2528
6	±2σ	0.4882	0.3452	0.2819	0.2441	0.2183	0.1993	0.1845	0.1726	0.1627	0.1544
	T.I.	0.8545	0.5822	0.4693	0.4037	0.3596	0.3274	0.3026	0.2826	0.2661	0.2522
7	±2σ	0.4880	0.3450	0.2817	0.2440	0.2182	0.1992	0.1844	0.1725	0.1627	0.1543
	T.I.	0.8453	0.5788	0.4673	0.4024	0.3586	0.3266	0.3019	0.2821	0.2657	0.2518
8	±2σ	0.4878	0.3449	0.2816	0.2439	0.2181	0.1991	0.1844	0.1725	0.1626	0.1542
	T.I.	0.8384	0.5761	0.4658	0.4014	0.3579	0.3260	0.3014	0.2816	0.2653	0.2515
9	±2σ	0.4876	0.3448	0.2815	0.2438	0.2181	0.1991	0.1843	0.1724	0.1625	0.1542
	T.I.	0.8330	0.5741	0.4646	0.4006	0.3573	0.3256	0.3010	0.2813	0.2650	0.2513
10	±2σ	0.4875	0.3447	0.2815	0.2438	0.2180	0.1990	0.1843	0.1724	0.1625	0.1542
	T.I.	0.8286	0.5724	0.4637	0.3999	0.3568	0.3252	0.3007	0.2811	0.2648	0.2511



ICH Q14 – Analytical Procedure Lifecycle





- Analytical Capability*
- Specificity
- Filter Validation
- Sample Solution Stability
- Accuracy*
- Linearity & Range
- Repeatability*
- Accuracy / Linearity / Repeatability*
[Combined as per ICH Q2(R1)]
- LOQ*, LOD*
- Intermediate Precision and Reproducibility
- Validation Robustness – LC
- Validation Robustness – Non-LC
[e.g. Sample Preparation, Dissolution]
- Method Transfer Study Support*

* – integration of USP <1210> Tolerance & Prediction Intervals]

Analytical Method Transfer Example

Transferring Lab



Fusion QbD
Sequence
Execution

Chromatography
Data Software

ALR
Design




Fusion QbD
Sequence
Execution

Chromatogram
Results Data

Receiving Lab




1. Fusion QbD – Exports experiment to the CDS as Ready-to-Run sequence, methods, standards
2. Sequence is run at both labs.
3. Fusion QbD – Imports results for instant and complete analysis and reporting.



Accuracy
Linearity
Repeatability
Tolerance Interval
Pass/Fail Results

ICH Q14 – Analytical Procedure Lifecycle



Analytical Control Strategy



ICH Q14

Knowledge gained from application of an enhanced approach to analytical procedure development can provide better assurance of the performance of the procedure, can serve as a basis for the analytical procedure control strategy and can provide an opportunity for more efficient regulatory approaches to related post approval changes.

USP <1220>


The [Analytical Control Strategy] is a set of controls needed to ensure the procedure performs as expected and plays a key role in ensuring that the ATP is realized throughout the life cycle. The preliminary ACS is identified during the procedure development process in Stage 1, ...

Variable Settings			
Enabled	Experiment Variable	Units	Maximum Expected Variation ($\pm 3\sigma$ Value)
<input checked="" type="checkbox"/>	Pump Flow Rate	mL/min	0.020
<input checked="" type="checkbox"/>	Oven Temperature	$^{\circ}\text{C}$	3.0
<input checked="" type="checkbox"/>	pH	*	0.15
<input checked="" type="checkbox"/>	Mobile Phase Composition (MPC)*	%	2.0

* - MPC variation is composition (blend) variation due to pump precision limits. A commonly used $\pm 3\sigma$ value = $\pm 2.0\%$.
The value you enter will be applied to all Gradient Slope factors (e.g., Time, Slope, and Ramp Steps) in the experiment design.



LC System Control Specifications



	Name	Units	Goal	Lower Bound	Upper Bound	Color
<input checked="" type="checkbox"/>	A - ResolutionW50	*	Maximize	2.000		Red
<input checked="" type="checkbox"/>	API - ResolutionW50	*	Maximize	2.000		Blue
<input checked="" type="checkbox"/>	D-Deg - ResolutionW50	*	Maximize	2.000		Green
<input checked="" type="checkbox"/>	E - ResolutionW50	*	Maximize	2.000		Orange
<input checked="" type="checkbox"/>	B - RetentionTime		Maximize	1.00		Gray
<input checked="" type="checkbox"/>	API - USPTailing		Minimize		1.50	Purple
<input checked="" type="checkbox"/>	B - RetentionTime - Cpk	*	Maximize	1.330		Gray
<input checked="" type="checkbox"/>	API - USPTailing - Cpk	*	Maximize	1.330		Purple
<input checked="" type="checkbox"/>	A - ResolutionW50 - Cpk	*	Maximize	1.330		Red
<input checked="" type="checkbox"/>	API - ResolutionW50 - Cpk	*	Maximize	1.330		Blue
<input checked="" type="checkbox"/>	D-Deg - ResolutionW50 - Cpk	*	Maximize	1.330		Green
<input checked="" type="checkbox"/>	E - ResolutionW50 - Cpk	*	Maximize	1.330		Orange



Routine Monitoring – Control Charts

